



# PAR Quantum Sensor SKP 215

Skye Instruments have been specialising in light and radiation sensors since 1983. All are designed, manufactured and calibrated to the highest standards. Each is supplied with a Calibration Certificate traceable to the UK's National Physical Laboratory (NPL).

There are three PAR sensors in the range, PAR Quantum, PAR Special and PAR Energy models. All measure the Photosynthetically Active Radiation between 400-700nm, the part of the solar spectrum used by plants for photosynthesis and sugar production.

The most popular is the PAR Quantum sensor which is used to measure photon irradiance, or quantity of PAR light. It is calibrated in units of  $\mu\text{mol m}^{-2} \text{s}^{-1}$  (number or quanta of photons). Sensors are suitable for use in natural solar radiation or other lamps or light sources. Each is fully weatherproof.

As with all Skye sensors, the PAR Quantum sensor has been quoted in many scientific references, please ask for a list of publications. They are compatible with Skye Display Meters, SpectroSense2 meters and DataHog2 dataloggers. A choice of outputs are also available to suit most dataloggers and controllers.



Measures Photosynthetically Active Radiation  
Counts quanta of photons in  $\mu\text{mol m}^{-2} \text{s}^{-1}$   
Ideal or square PAR spectrum response  
For plant and crop research  
Commercial horticulture applications  
Suitable for natural and artificial light sources

## SKP 215 SPECIFICATIONS

**Construction** - Material Dupont 'Delrin' fully sealed to Ip67

**Cable** - Screened 7-2-3C

**Sensor** - Cosine corrected head

**Detector** - Blue enhanced planar diffused silicon

**Filters** - Optical glass

**Sensitivity - current (1)** -  $0.015\mu\text{A}/\mu\text{mol m}^{-2} \text{s}^{-1}$

**Sensitivity - voltage** -  $10\mu\text{V}/\mu\text{mol m}^{-2} \text{s}^{-1}$

**Working range (2)** -  $0-5 \times 10^4 \mu\text{mol m}^{-2} \text{s}^{-1}$

**Linearity error** -  $<0.2\%$

**Absolute calibration error (3)** - typ.  $<3\%$ , 5% max

**Cosine error (4)** - 3%

**Azimuth error (5)** -  $<1\%$

**Temperature coefficient** -  $\pm 0.1\%/^{\circ}\text{C}$

**Longterm stability (6)** -  $\pm 2\%$

**Response time (7) (voltage output)** - 10ns

**Internal resistance (voltage output)** - c.300 ohms

**Temperature range** -  $-35$  to  $+75^{\circ}\text{C}$

**Humidity range** - 0-100% RH

**Weight** - 130g (with 3m cable)

**Dimensions** -



## ORDERING INFORMATION

### Sensor

SKP 215 - PAR Quantum Sensor with 3m cable

### Accessories

SKM 221 - Levelling unit

SKM 226 - Long arm pole/wall mount

### Meters and Dataloggers

SKP 200 - Display meter

SKL 904 - 4 channel SpectroSense2 display meter

SKL 908 - 8 channel SpectroSense2 logging display meter

SDL 5000 series DataHog datalogger

## NOTES ON SPECIFICATIONS

(1) Current output varies from sensor to sensor. Each individual unit will have a slightly different output. A calibration certificate is supplied with each sensor.

(2) All Skye sensors will work at levels of irradiance well above that found in terrestrial sunlight conditions, room or growth chamber lighting.

(3) Main source of this error is uncertainty of calibration of Reference Lamp. Skye calibration standards are directly traceable to N.P.L. Standard references.

(4) Cosine error to  $80^{\circ}$  is typically 5% max. Figures shown are for normal use sources, e.g., sun plus sky, diffuse sun, growth chambers, etc.

(5) Measured at  $45^{\circ}$  elevation over  $360^{\circ}$ .

(6) Maximum change in one year. Calibration check recommended at least every two years. Experience has shown that changes are typically much less than figures quoted.

(7) Times are generally less than the figure quoted, which is in nanoseconds. They may be slightly increased if long leads are fitted, or those of a higher capacity cable.

## GRAPH

### PAR Quantum - SKP 215

