



Technical Help WindSonic

WindSonic - 4-20mA Output

PLEASE REFER TO PAGE 24 OF THE WINDSONIC USER MANUAL FOR WIRING DETAILS, ALONG WITH THE CONNECTOR WIRING BELOW. ALSO REFER TO THE WINDSONIC USER MANUAL DOC NO.1405-PS-0019, ISSUE 16 OR LATER FOR UNITS WITH A SERIAL NUMBER 08100001 ONWARDS.

The following details are for use with WindSonic OPTION 3 and are the default settings that Skye Instruments will set unless otherwise requested at the time of ordering.

Setting up the WindSonic.

With reference to page 42 of the WindSonic manual, enter configuration mode, then check the configuration. It may be a good idea to record the current settings. The following settings are required:

Measurement settings	M2	Gill, Polar, Continuous
Output units	U1	Meters per second
ASCII output format	O1 or O2	
Message terminator	L1 or L2	
Output rate	P1	1 per second
Baud rate	B3	9600
Data and parity options	F1	8 bits, no parity
Power-up message	H2	Off
Node address	NQ	
Analogue settings	T3	4-20mA
	S7	0 to 60 metres per second
	C2	Normal 0 359 degrees

WindSonic 9-way circular connector wiring

Pin Number	Wire Colour	Function
1	Brown	Ground
2	Red	Positive power supply (7-30 VDC)
3	Orange	Negative power supply
4	Yellow	See manual
5	Green	RS232 TXD
6	Blue	See manual
7	Purple	RS232 RXD
8	Grey	Wind speed output (positive current relative to Pin 1)
9	White	Wind direction output (positive current relative to Pin 1)

The screen of the cable is left uncommitted.

Skye Instruments Ltd

21, Ddole Enterprise Park, Llandrindod Wells
Powys LD1 6DF, United Kingdom

TEL +44 (0)1597 824811
FAX +44 (0)1597 824812

EMAIL skyeemail@skyeinstruments.com
WEB <http://www.skyeinstruments.com>